Natrelle® INSPIRA® COLLECTION

Natrelle® 133S smooth tissue expander

Natrelle® 133S Smooth Tissue Expanders

Select an expander below based on the width measurement and find the complementary range of *Natrelle* INSPIRA® Smooth Round Gel Breast Implants to the right in the same color.*

Full Height, Variable Projection
Style FV-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	11.5	5.0	300	133S-FV-11-T
12.0	12.5	5.3	400	133S-FV-12-T
13.0	13.5	5.7	500	133S-FV-13-T
14.0	14.5	6.2	600	133S-FV-14-T
15.0	15.5	6.7	750	133S-FV-15-T
16.0	16.5	6.8	850	133S-FV-16-T

Moderate Height, Variable Projection
Style MV-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	10.0	4.9	250	133S-MV-11-T
12.0	11.0	5.2	300	133S-MV-12-T
13.0	12.0	5.6	400	133S-MV-13-T
14.0	13.0	6.0	500	133S-MV-14-T
15.0	14.0	6.3	600	133S-MV-15-T
16.0	15.0	6.6	700	133S-MV-16-T

Short Height, Variable Projection
Style SV-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	7.6	4.7	200	133S-SV-11-T
12.0	8.4	5.3	250	133S-SV-12-T
13.0	9.1	5.7	300	133S-SV-13-T
14.0	9.7	6.0	375	133S-SV-14-T
15.0	10.5	6.3	450	133S-SV-15-T
16.0	11.3	6.3	550	133S-SV-16-T

Note: *Natrelle*® 133S Smooth Tissue Expanders have 6 suture tabs (except for Style SV, which has 4).

*For implant widths from 11.0 cm to 16.5 cm.

Full Height, Extra Projection
Style FX-T*

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	11.5	5.9	350	133S-FX-11-T
12.0	12.5	6.3	450	133S-FX-12-T
13.0	13.5	6.7	550	133S-FX-13-T
14.0	14.5	7.1	650	133S-FX-14-T
15.0	15.5	7.6	800	133S-FX-15-T
15.5	16.0	8.1	950	133S-FX-16-T

Moderate Height, Extra Projection
Style MX-T*

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	10.0	5.9	300	133S-MX-11-T
12.0	11.0	6.3	400	133S-MX-12-T
13.0	12.0	6.7	500	133S-MX-13-T
14.0	13.0	7.1	600	133S-MX-14-T
15.0	14.0	7.6	700	133S-MX-15-T
15.5	14.5	8.1	850	133S-MX-16-T

Short Height, Extra Projection
Style SX-T*

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	9.0	5.9	250	133S-SX-11-T
12.0	10.0	6.3	350	133S-SX-12-T
13.0	11.0	6.7	400	133S-SX-13-T
14.0	12.0	7.1	500	133S-SX-14-T
15.0	13.0	7.6	650	133S-SX-15-T
15.5	13.5	8.1	800	133S-SX-16-T

PRO TIP:
Use the color system to guide implant selection. Choose an implant that is up to 1.0 cm wider than the patient's breast base width to help ensure an optimal fit in the breast pocket.

Natrelle INSPIRA® Style SRL, SSL, SCL
Smooth, Low Profile



WIDTH (CM)	PROJECTION (CM)	VOLUME (CC)	SMOOTH RESPONSIVE (SRL)	SMOOTH SOFTTOUCH (SSL)	SMOOTH COHESIVE (SCL)
---	---	---	---	---	---
---	---	---	---	---	---
10.00	2.0	110	SRL-110	SSL-110	SCL-110
10.25	2.0	125	SRL-125	SSL-125	SCL-125
10.50	2.1	140	SRL-140	SSL-140	SCL-140
---	---	---	---	---	---
11.00	2.2	170	SRL-170	SSL-170	SCL-170
---	---	---	---	---	---
---	---	---	---	---	---
11.50	2.3	200	SRL-200	SSL-200	SCL-200
---	---	---	---	---	---
12.00	2.4	230	SRL-230	SSL-230	SCL-230
---	---	---	---	---	---
12.50	2.5	260	SRL-260	SSL-260	SCL-260
---	---	---	---	---	---
13.00	2.6	290	SRL-290	SSL-290	SCL-290
---	---	---	---	---	---
13.50	2.7	320	SRL-320	SSL-320	SCL-320
---	---	---	---	---	---
14.00	2.8	350	SRL-350	SSL-350	SCL-350
---	---	---	---	---	---
14.50	2.9	380	SRL-380	SSL-380	SCL-380
---	---	---	---	---	---
15.00	3.0	410	SRL-410	SSL-410	SCL-410
15.50	3.1	460	SRL-460	SSL-460	SCL-460
16.00	3.2	510	SRL-510	SSL-510	SCL-510
16.50	3.3	560	SRL-560	SSL-560	SCL-560
17.00	3.4	610	SRL-610	SSL-610	SCL-610

Natrelle INSPIRA® Style SRLP, SSLP, SCLP
Smooth, Low-Plus Profile



WIDTH (CM)	PROJECTION (CM)	VOLUME (CC)	SMOOTH RESPONSIVE (SRLP)	SMOOTH SOFTTOUCH (SSLP)	SMOOTH COHESIVE (SCLP)
9.50	2.5	125	SRLP-125	SSLP-125	SCLP-125
---	---	---	---	---	---
10.00	2.6	145	SRLP-145	SSLP-145	SCLP-145
---	---	---	---	---	---
10.50	2.7	165	SRLP-165	SSLP-165	SCLP-165
---	---	---	---	---	---
11.00	2.8	190	SRLP-190	SSLP-190	SCLP-190
11.25	2.9	205	SRLP-205	SSLP-205	SCLP-205
---	---	---	---	---	---
11.50	3.0	220	SRLP-220	SSLP-220	SCLP-220
11.75	3.1	235	SRLP-235	SSLP-235	SCLP-235
12.00	3.1	250	SRLP-250	SSLP-250	SCLP-250
12.25	3.2	265	SRLP-265	SSLP-265	SCLP-265
12.50	3.3	280	SRLP-280	SSLP-280	SCLP-280
12.75	3.3	300	SRLP-300	SSLP-300	SCLP-300
13.00	3.4	320	SRLP-320	SSLP-320	SCLP-320
13.25	3.5	340	SRLP-340	SSLP-340	SCLP-340
13.50	3.5	360	SRLP-360	SSLP-360	SCLP-360
---	---	---	---	---	---
14.00	3.6	400	SRLP-400	SSLP-400	SCLP-400
---	---	---	---	---	---
14.50	3.8	440	SRLP-440	SSLP-440	SCLP-440
---	---	---	---	---	---
15.00	3.9	490	SRLP-490	SSLP-490	SCLP-490
15.50	4.0	540	SRLP-540	SSLP-540	SCLP-540
16.00	4.2	590	SRLP-590	SSLP-590	SCLP-590
16.50	4.3	640	SRLP-640	SSLP-640	SCLP-640
---	---	---	---	---	---

Natrelle INSPIRA® Style SRM, SSM, SCM
Smooth, Moderate Profile



WIDTH (CM)	PROJECTION (CM)	VOLUME (CC)	SMOOTH RESPONSIVE (SRM)	SMOOTH SOFTTOUCH (SSM)	SMOOTH COHESIVE (SCM)
9.50	3.0	140	SRM-140	SSM-140	SCM-140
---	---	---	---	---	---
10.00	3.3	175	SRM-175	SSM-175	SCM-175
10.25	3.4	195	SRM-195	SSM-195	SCM-195
10.50	3.5	210	SRM-210	SSM-210	SCM-210
---	---	---	---	---	---
11.00	3.7	240	SRM-240	SSM-240	SCM-240
11.25	3.8	255	SRM-255	SSM-255	SCM-255
---	---	---	---	---	---
11.50	3.9	275	SRM-275	SSM-275	SCM-275
11.75	3.9	295	SRM-295	SSM-295	SCM-295
12.00	4.0	310	SRM-310	SSM-310	SCM-310
12.25	4.1	330	SRM-330	SSM-330	SCM-330
12.50	4.2	345	SRM-345	SSM-345	SCM-345
12.75	4.2	360	SRM-360	SSM-360	SCM-360
13.00	4.3	375	SRM-375	SSM-375	SCM-375
13.25	4.4	405	SRM-405	SSM-405	SCM-405
13.50	4.5	445	SRM-445	SSM-445	SCM-445
---	---	---	---	---	---
14.00	4.6	485	SRM-485	SSM-485	SCM-485
---	---	---	---	---	---
14.50	4.8	520	SRM-520	SSM-520	SCM-520
14.75	4.9	560	SRM-560	SSM-560	SCM-560
15.00	4.9	600	SRM-600	SSM-600	SCM-600
15.50	5.0	640	SRM-640	SSM-640	SCM-640
16.00	5.2	685	SRM-685	SSM-685	SCM-685
16.50	5.0	755	SRM-755	SSM-755	SCM-755
---	---	---	---	---	---

Natrelle INSPIRA® Style SRF, SSF, SCF
Smooth, Full Profile



WIDTH (CM)	PROJECTION (CM)	VOLUME (CC)	SMOOTH RESPONSIVE (SRF)	SMOOTH SOFTTOUCH (SSF)	SMOOTH COHESIVE (SCF)
9.50	4.0	180	SRF-180	SSF-180	SCF-180
9.75	4.1	200	SRF-200	SSF-200	SCF-200
10.00	4.2	220	SRF-220	SSF-220	SCF-220
---	---	---	---	---	---
10.50	4.3	240	SRF-240	SSF-240	SCF-240
---	---	---	---	---	---
11.00	4.4	265	SRF-265	SSF-265	SCF-265
11.25	4.5	295	SRF-295	SSF-295	SCF-295
11.40	4.6	325	SRF-325	SSF-325	SCF-325
11.50	4.7	335	SRF-335	SSF-335	SCF-335
11.75	4.8	345	SRF-345	SSF-345	SCF-345
12.00	4.9	365	SRF-365	SSF-365	SCF-365
12.25	5.0	385	SRF-385	SSF-385	SCF-385
12.50	5.1	415	SRF-415	SSF-415	SCF-415
12.75	5.3	450	SRF-450	SSF-450	SCF-450
13.00	5.4	485	SRF-485	SSF-485	SCF-485
13.25	5.5	520	SRF-520	SSF-520	SCF-520
13.50	5.7	560	SRF-560	SSF-560	SCF-560
---	---	---	---	---	---
14.00	5.8	605	SRF-605	SSF-605	SCF-605
---	---	---	---	---	---
14.50	5.9	650	SRF-650	SSF-650	SCF-650
14.75	6.0	695	SRF-695	SSF-695	SCF-695
15.00	6.1	745	SRF-745	SSF-745	SCF-745
15.50	6.0	770	SRF-770	SSF-770	SCF-770
---	---	---	---	---	---
---	---	---	---	---	---
---	---	---	---	---	---
---	---	---	---	---	---

Natrelle INSPIRA® Style SRX, SSX, SCX
Smooth, Extra-Full Profile



WIDTH (CM)	PROJECTION (CM)	VOLUME (CC)	SMOOTH RESPONSIVE (SRX)	SMOOTH SOFTTOUCH (SSX)	SMOOTH COHESIVE (SCX)
9.50	4.2	200	SRX-200	SSX-200	SCX-200
---	---	---	---	---	---
10.00	4.5	225	SRX-225	SSX-225	SCX-225
10.25	4.8	255	SRX-255	SSX-255	SCX-255
10.50	5.0	285	SRX-285	SSX-285	SCX-285
10.75	5.2	310	SRX-310	SSX-310	SCX-310
11.00	5.4	340	SRX-340	SSX-340	SCX-340
11.25	5.6	375	SRX-375	SSX-375	SCX-375
---	---	---	---	---	---
11.50	5.8	400	SRX-400	SSX-400	SCX-400
11.75	5.9	420	SRX-420	SSX-420	SCX-420
12.00	6.0	445	SRX-445	SSX-445	SCX-445
12.25	6.1	470	SRX-470	SSX-470	SCX-470
12.50	6.1	495	SRX-495	SSX-495	SCX-495
12.75	6.2	525	SRX-525	SSX-525	SCX-525
13.00	6.2	545	SRX-545	SSX-545	SCX-545
13.25	6.3	560	SRX-560	SSX-560	SCX-560
13.50	6.4	580	SRX-580	SSX-580	SCX-580
13.75	6.4	615	SRX-615	SSX-615	SCX-615
14.00	6.5	650	SRX-650	SSX-650	SCX-650
14.25	6.6	700	SRX-700	SSX-700	SCX-700
14.50	6.7	750	SRX-750	SSX-750	SCX-750
---	---	---	---	---	---
---	---	---	---	---	---
15.50	6.7	800	SRX-800	SSX-800	SCX-800
---	---	---	---	---	---
---	---	---	---	---	---
---	---	---	---	---	---

Natrelle® Breast Implants
IMPORTANT SAFETY INFORMATION

WARNINGS
- Breast implants are not considered lifetime devices. The longer patients have them, the greater the chance they will develop complications, some of which will require more surgery
- Breast implants have been associated with the development of a cancer of the immune system called breast implant–associated anaplastic large cell lymphoma (BIA-ALCL). This cancer occurs more commonly in patients with textured breast implants than smooth implants, although rates are not well defined. Some patients have died from BIA-ALCL
- Patients receiving breast implants have reported a variety of systemic symptoms, such as joint pain, muscle aches, confusion, chronic fatigue, autoimmune diseases, and others. Individual patient risk for developing these symptoms has not been well established. Some patients report complete resolution of symptoms when the implants are removed without replacement

Please see additional Important Safety Information on back.

INDICATIONS
Natrelle® Breast Implants are indicated for women for the following:
- **Breast augmentation for women at least 22 years old for silicone-filled implants and breast augmentation for women at least 18 years old for saline-filled implants.** This includes primary breast augmentation to increase the breast size, as well as revision surgery to correct or improve the result of a primary breast augmentation surgery
- **Breast reconstruction.** This includes primary reconstruction to replace breast tissue that has been removed due to cancer or trauma or that has failed to develop properly due to a severe breast abnormality. Breast reconstruction also includes revision surgery to correct or improve the result of a primary breast reconstruction surgery

CONTRAINDICATIONS
Breast implant surgery should not be performed in:
- Women with active infection anywhere in their body
- Women with existing cancer or precancer of their breast who have not received adequate treatment for those conditions
- Women who are currently pregnant or nursing

ADDITIONAL WARNINGS
- See Boxed Warning
- **Avoid damage during surgery:** Care should be taken to avoid the use of excessive force and to minimize handling of the implant. Forcing of implants through too small an opening or applying concentrated localized pressure on the implants may result in localized weakening of the breast implant shell, potentially leading to shell damage and possible implant rupture. An incision should be of appropriate length to accommodate the style, size, and profile of the implants. Use care when using surgical instruments in proximity with the breast implant
- Follow recommended fill volumes for saline implants to decrease possibility of shell wrinkling and crease-fold failure

PRECAUTIONS
Safety and effectiveness have not been established in patients with the following:
- Autoimmune diseases (eg, lupus and scleroderma)
- A compromised immune system (eg, currently receiving immunosuppressive therapy)
- Planned chemotherapy or radiation following breast implant placement
- Conditions or medications that interfere with wound healing and blood clotting
- Reduced blood supply to breast tissue
- Clinical diagnosis of depression or other mental health disorders, including body dysmorphic disorder and eating disorders. Please discuss any history of mental health disorders prior to surgery. Patients with a diagnosis of depression, or other mental health disorders, should wait until resolution or stabilization of these conditions prior to undergoing breast implantation surgery

ADVERSE EVENTS
Possible adverse events with breast implant surgery include implant rupture with silicone implants, implant deflation with saline-filled implants, capsular contracture, reoperation, implant removal, pain, changes in nipple and breast sensation, infection, scarring, asymmetry, wrinkling, implant displacement/migration, implant palpability/visibility, breastfeeding complications, hematoma/seroma, implant extrusion, necrosis, delayed wound healing, infection, breast tissue atrophy/chest wall deformity, calcium deposits, and lymphadenopathy. Other systemic conditions have been reported with breast implants.

For more information, please see the full Directions for Use at www.allergan.com/products.

To report a problem with *Natrelle*® Breast Implants, please call Allergan® at 1-800-624-4261.

The sale and distribution of this device is restricted to users and/or user facilities that provide information to patients about the risks and benefits of this device in the form and manner specified in the approved labeling provided by Allergan®.

What's in a SKU?
Understand your options for easy ordering



For instance:
- **SCM-310** means Smooth surface, Cohesive gel, Moderate profile, 310 cc volume
- **SSLP-250** means Smooth surface, SoftTouch gel, Low-Plus profile, 250 cc volume

MATCHUP GUIDE

Match *Natrelle*® 133S Smooth Tissue Expanders With *Natrelle* INSPIRA® Responsive, Cohesive, and SoftTouch Round Gel Breast Implants.

***Natrelle* INSPIRA® Smooth Round Gel Breast Implants**



***Natrelle*® 133S Smooth Tissue Expander**



FOURTÉ® Expander Fill System
Included with all *Natrelle*® 133S Smooth Tissue Expanders

Natrelle®

Please see Important Safety Information inside, including Boxed Warning.

Navigating the *Natrelle*® 133S Smooth Tissue Expander System: A Biodimensional Planning Guide

Not all breast tissue expanders *are created equal*

The *Natrelle*® 133S Smooth Tissue Expander System was developed with biodimensional planning in mind. Our smooth tissue expanders are available in 36 options and 6 shapes across 2 profiles to meet the unique needs of each patient.[1]

Paired for precision *for the breast shape you envision*[2]





100% MATCH

between all *Natrelle*® 133S Smooth Tissue Expanders and *Natrelle* INSPIRA® breast implants (in both base width *and* projection).[2,4]

[1]**Methodology** Tissue expanders were matched to round breast implants. A precise match was defined as a breast implant that falls within a range of 0.5 cm smaller to 1.0 cm larger in base width and 1.5 cm smaller to 2.0 cm larger in projection of a tissue expander. The percentage of precise matches was calculated for each style of tissue expander.

Why is biodimensional planning important?

Patient anatomy is important in tissue expander selection. Dimensions should be consistent with the patient in terms of[3,4]:

- Breast base width
- Breast height
- Chest wall anatomy

STEP 1

MEASURE the *true base width* of the breast

Accurately measuring the true base width of the patient's breast is key to determining which *Natrelle*® 133S Smooth Tissue Expander to use.



Measure the base width of the breast in centimeters using either calipers or a tape measure.

STEP 2

MEASURE the *height* of the breast

It is also important to measure the patient's breast height. Be aware of variations in patient anatomy when taking this measurement.



Measure the height of the breast in centimeters using either calipers or a tape measure.

STEP 3

USE MEASUREMENTS to guide *tissue expander selection*

The true base width and height of the breast are the key measurements that help determine the best tissue expander to use for each patient from the *Natrelle*® 133S Smooth Tissue Expander portfolio.

Full Height, Variable Projection
Style FV-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	11.5	5.0	300	133S-FV-11-T
12.0	12.5	5.3	400	133S-FV-12-T
13.0	13.5	5.7	500	133S-FV-13-T
14.0	14.5	6.2	600	133S-FV-14-T
15.0	15.5	6.7	750	133S-FV-15-T
16.0	16.5	6.8	850	133S-FV-16-T

Moderate Height, Variable Projection
Style MV-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	10.0	4.9	250	133S-MV-11-T
12.0	11.0	5.2	300	133S-MV-12-T
13.0	12.0	5.6	400	133S-MV-13-T
14.0	13.0	6.0	500	133S-MV-14-T
15.0	14.0	6.3	600	133S-MV-15-T
16.0	15.0	6.6	700	133S-MV-16-T

Short Height, Variable Projection
Style SV-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	7.6	4.7	200	133S-SV-11-T
12.0	8.4	5.3	250	133S-SV-12-T
13.0	9.1	5.7	300	133S-SV-13-T
14.0	9.7	6.0	375	133S-SV-14-T
15.0	10.5	6.3	450	133S-SV-15-T
16.0	11.3	6.3	550	133S-SV-16-T

WIDTH (CM)	HEIGHT (CM)
12.0	12.5

Choose a tissue expander that is 0.5 to 1.0 cm narrower than the base width measurement. Locate this and the breast height measurement in the first 2 columns of the tissue expander portfolio chart.

Example: The true base width of your patient's breast measures 12.5 cm and the breast height measures 12.5 cm. Based on this, you would select an expander with a 12.0 cm width and 12.5 cm height.

Variable Projection

Extra Projection

Chest wall concavity or convexity can help determine whether a variable or extra projection tissue expander may be the patient's best option.

PRO TIP:
If a patient's measurements fall in between two numbers, choose the smaller tissue expander to help ensure a tighter implant pocket.

Full Height, Extra Projection
Style FX-T[1]

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	11.5	5.9	350	133S-FX-11-T
12.0	12.5	6.3	450	133S-FX-12-T
13.0	13.5	6.7	550	133S-FX-13-T
14.0	14.5	7.1	650	133S-FX-14-T
15.0	15.5	7.6	800	133S-FX-15-T
15.5	16.0	8.1	950	133S-FX-16-T

Moderate Height, Extra Projection
Style MX-T

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	10.0	5.9	300	133S-MX-11-T
12.0	11.0	6.3	400	133S-MX-12-T
13.0	12.0	6.7	500	133S-MX-13-T
14.0	13.0	7.1	600	133S-MX-14-T
15.0	14.0	7.6	700	133S-MX-15-T
15.5	14.5	8.1	850	133S-MX-16-T

Short Height, Extra Projection
Style SX-T[1]

WIDTH (CM)	HEIGHT (CM)	PROJECTION (CM)	FILL VOLUME (CC)	CATALOG NUMBER
11.0	9.0	5.9	250	133S-SX-11-T
12.0	10.0	6.3	350	133S-SX-12-T
13.0	11.0	6.7	400	133S-SX-13-T
14.0	12.0	7.1	500	133S-SX-14-T
15.0	13.0	7.6	650	133S-SX-15-T
15.5	13.5	8.1	800	133S-SX-16-T

Natrelle® 133S Smooth Tissue Expanders With MAGNA-SITE® Injection Sites IMPORTANT SAFETY INFORMATION

INDICATIONS

Natrelle® 133S Smooth Tissue Expanders are indicated for:

- Breast reconstruction following mastectomy
- Treatment of underdeveloped breasts
- Treatment of soft tissue deformities

IMPORTANT SAFETY INFORMATION

CONTRAINDICATIONS

Natrelle® 133S Smooth Tissue Expanders **should not** be used in patients:

- Who already have implanted devices that would be affected by a magnetic field (eg, pacemakers, drug infusion devices, and artificial sensing devices)
- Whose tissue at the expansion site is determined to be unsuitable
- Who have an active infection or a residual gross tumor at the expansion site
- Who are undergoing adjuvant radiation therapy
- Whose physiological condition (eg, sensitive over- or underlying anatomy, obesity, smoking, diabetes, autoimmune disease, hypertension, chronic lung or severe cardiovascular disease, or osteogenesis imperfecta) or use of certain drugs (including those that interfere with blood clotting or affect tissue viability) poses an unduly high risk of surgical and/or postoperative complications
- Who are psychologically unsuitable

WARNINGS

- **Do not** use *Natrelle*® 133S Smooth Tissue Expanders in patients who already have implanted devices that would be affected by a magnetic field (see *Contraindications*) because the MAGNA-SITE® integrated injection site contains a strong rare-earth, permanent magnet. Diagnostic testing with Magnetic Resonance Imaging (MRI) is contraindicated in patients with *Natrelle*® 133S Smooth Tissue Expanders in place
- **Do not** alter the tissue expander or use adulterated fill. Fill only with sterile saline for injection as described in INSTRUCTIONS FOR USE. **Do not** expose to contaminants
- **Do not** expand if the pressure will compromise wound healing or vasculature of overlying tissue, or beyond patient or tissue tolerance. Stop filling immediately if tissue damage, wound dehiscence, abnormal skin pallor, erythema, edema, pain, or tenderness are observed
- **Do not reuse explanted products**
- Active infection anywhere may increase risk of periprosthetic infection. **Do not expose the tissue expander or injection needles to contaminants.** Postoperative infections should be treated aggressively
- Adverse reactions may require premature explantation
- When using suturing tabs be careful to avoid piercing the shell. Use a new one if damage occurs
- ***Natrelle*® 133S Smooth Tissue Expanders are temporary devices and are not to be used for permanent implantation or beyond 6 months.** Tissue expansion in breast reconstruction typically requires 4 to 6 months

PRECAUTIONS

Active infections may need to be treated and resolved before surgery. Follow proper surgical procedures and carefully evaluate patient suitability using standard practice and individual experience. Avoid damage to the tissue expander and use a **sterile backup** in case of damage. Pay careful attention to tissue tolerance and hemostasis during surgery. Expansion should proceed moderately and never beyond patient or tissue tolerance. Avoid contamination in any postoperative procedure.

ADVERSE REACTIONS

Possible adverse reactions include deflation, tissue damage, infection, extrusion, hematoma/seroma, capsular contracture, premature explantation, displacement, effects on bone, pain, sensation, distortion, inadequate tissue flap, and inflammatory reaction.

For more information, please see the full Directions for Use at www.allergan.com/products.

To report a problem with *Natrelle*® 133S Smooth Tissue Expanders, please call Allergan® at 1-800-624-4261.

Natrelle® 133S Smooth Tissue Expanders are restricted to sale by or on the order of a licensed physician.

References: 1. Data on file, Allergan Aesthetics, Hospital Order Form Update, January 2022. 2. Data on file, Allergan Aesthetics, January 6, 2017; Study Report MD16076-DV. 3. Nava MB, Rocco N, Tunesi G, Catanuto G, Rancati A, Dorr J. Decisional pathways in breast augmentation: how to improve outcomes through accurate pre-operative planning. *Gland Surg*. 2017;6(2):203-209. 4. Movassaghi K, Cusic J. Shaping the breast: Optimizing outcomes in breast augmentation. In: Movassaghi K, ed. *Shaping the Breast. A Comprehensive Approach in Augmentation, Revision, and Reconstruction*. Springer; 2021:1-34.

Allergan Aesthetics
an AbbVie company

© 2024 AbbVie. All rights reserved. NATRELLE and its design are trademarks of Allergan, Inc., an AbbVie company. Natrelle.com NAT604569-v7 11/24 025825